

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 28 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130173</u>	
Name <u>J Hembough</u>		ID# <u>5828</u>	Date <u>4-26-10</u>
A	Agency <u>AST</u>	Phone # <u>800.3263</u>	
Instrument Location <u>Glennallen</u>			
Alco S/N <u>X 124669</u>		Target Value <u>0.083</u>	High Pressure <u>1100</u>
B	Alco Test Values <u>0.080</u> 1 st Alco <u>0.080</u> 2 nd Alco		
Signature <u>[Signature]</u>			
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 8th day of June, 2010.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130173

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130173

APRIL 25, 2010

OPERATOR'S NAME:

HEINBAUGH

OPERATOR'S NUMBER: 5828

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: GAKO

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: 0.083

ALCO S/N: X124669

— BREATH ANALYSIS —

.083 ADJUSTED FOR 28.24 in		
ALCO TARGET	.078	16:56
BLANK TEST	.000	16:57
INTERNAL STANDARD	VERIFIED	16:57
ALCO TV 28.22 in	.000	16:58
BLANK TEST	.000	16:58
SUBJECT SAMPLE	.000	16:59
BLANK TEST	.000	17:00
ALCO TV 28.24 in	.000	17:00
BLANK TEST	.000	17:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130173

APRIL 25, 2010

TIME 17:01

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 28.24 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~